

BLOOD GROUPING REAGENT

Anti-k̄ (cellano)

ALBAclone®

(Murine Monoclonal IgM)

For Tube Technique

This insert refers to product Z137U

- Meets FDA potency requirements
- Discard if turbid
- Preservative: 0.1% sodium azide

CAUTIONS: THE ABSENCE OF ALL VIRUSES HAS NOT BEEN DETERMINED. THIS PRODUCT HAS COMPONENTS (DROPPER BULBS) CONTAINING DRY NATURAL RUBBER.

SUMMARY

Since the description of the antigen K in 1946 by Coombs *et al* and its allele k in 1949 by Levine *et al*, the Kell blood group system has been shown to be increasingly complex. Over 20 antigens are now known to be associated with the system and 4 sets of alleles have been identified ie K, k; Kp^a, Kp^b, Kp^c; Js^a, Js^b; K11 (Cste) and K17 (Wk^a). These are probably controlled from a series of closely linked loci so that Kell antigens, like CDE in the Rh system, are inherited as a haplotype.

The antigens of the Kell blood group system are of further interest in that they tend to occur either very frequently (e.g. k 99.8%) or relatively infrequently (e.g. K 8%) and show considerable ethnic variation e.g. the antigen Js^a is extremely rare in whites but is expressed by 20% of black Americans. The antigens require the presence of disulphide bonds. Treatment with certain proteolytic enzymes may reduce reactivity.

Kell system antibodies are capable of causing haemolytic transfusion reactions and haemolytic disease of the newborn and are optimally detected by the indirect antiglobulin technique.

INTENDED USE

This Anti-k reagent is for the *in vitro* detection and identification of the human k blood group antigen by direct agglutination.

PRINCIPLE OF THE TEST

When used by the recommended technique, this reagent will cause agglutination (clumping) of red blood cells carrying the k (cellano) antigen. Lack of agglutination demonstrates the absence of the k (cellano) antigen.

REAGENT DESCRIPTION

The main component of this reagent is derived from the *in vitro* culture of the IgM secreting mouse hybridoma LKL1.

The formulation consists of culture supernatant in MES buffer pH 5.2 and 1g/l sodium azide.

The volume delivered by the reagent dropper bottle is approximately 40µl. Bearing this in mind, care should be taken to ensure that appropriate serum:cell ratios are maintained in all test systems.

STORAGE CONDITIONS

The reagent should be stored at 2°C - 8°C. Do not use if turbid. Do not dilute. The reagent is stable until the expiry date stated on the product label.

PRECAUTIONS FOR USE AND DISPOSAL

This reagent contains 0.1% (w/v) sodium azide. Sodium azide may be toxic if ingested and may react with lead and copper plumbing to form explosive compounds. If discarded into sink, flush with a large volume of water to prevent azide buildup.

As this reagent is of animal origin care must be taken during use and disposal as there is a potential infection risk.

This product has components (dropper bulbs) containing dry natural rubber. This reagent is for *in vitro* diagnostic use only.

SPECIMEN COLLECTION AND PREPARATION

Specimens should be collected by aseptic technique with or without an anticoagulant. The specimen should be tested as soon as possible after collection. If testing is delayed, the specimen should be stored at 2°C - 8°C. Blood specimens exhibiting gross haemolysis or contamination should not be used. Clotted samples or those collected in EDTA should be tested within fourteen days from collection. Donor blood stored in citrate anticoagulant may be tested until the expiry date of the donation.

TEST PROCEDURES

General Information

This reagent has been standardised for use by the technique described below and therefore its suitability for use in other techniques cannot be guaranteed. When a test is required to be incubated for a specific time period, a timer should be used.

ADDITIONAL MATERIALS AND REAGENTS REQUIRED

- Isotonic saline
- Reagent red blood cells suitable for the control of Anti-k
- 10 x 75mm or 12 x 75mm glass test tubes
- Pipettes
- Centrifuge
- Incubator
- Timer

RECOMMENDED TECHNIQUE

Tube Technique - 5 Minute Incubation / Spin

- Add 1 volume of blood grouping reagent to a test tube.
- Add 1 volume of red blood cells suspended to 2-4% in isotonic saline.
- Mix the contents of the test tube well and incubate at 20° - 25°C for 5 minutes.
- Centrifuge the test tube.
Suggested centrifugation: 1000g for 10 seconds or a time and speed appropriate for the centrifuge used that produces the strongest reaction of antibody with antigen-positive red blood cells, yet allows easy resuspension of antigen-negative red blood cells.
- After centrifugation, gently shake the tube to dislodge the cell button from the bottom and immediately observe macroscopically for agglutination.


INTERPRETATION OF LABELING SYMBOLS



Batch code



Use by (YYYY-MM-DD)



Storage temperature limitation (2°C-8°C)



In vitro diagnostic medical device



Consult instructions for use



Harmful



Manufacturer

INTERPRETATION OF RESULTS

Agglutination = positive test result
No agglutination = negative test result

QUALITY CONTROL

Quality control of reagents is essential and should be performed with each series of groups, single groups and in accordance with local, state and federal regulations. We suggest that the following red blood cell samples are used to control the reactions of this reagent.

- Kk red blood cells should be used as a positive control
- KK red blood cells should be used as a negative control

PERFORMANCE LIMITATIONS

Red blood cells from individuals of the Kell phenotype KkKp (a+b+) show a substantially weakened expression of k antigen.

Kell antigen expression may be dramatically weakened in some cases of Chronic Granulomatous Disease.

Gently resuspend tube tests before reading. Excessive agitation may disrupt weak agglutination and produce false negative results.

Excessive centrifugation can lead to difficulty in resuspending the cell button, while inadequate centrifugation may result in agglutinates that are easily dispersed.

The expression of certain red blood cell antigens may diminish in strength during storage, particularly in EDTA and clotted samples. Better results will be obtained with fresh samples.

False positive or false negative results can occur due to contamination of test materials, improper reaction temperature, improper storage of materials, omission of test reagents and certain disease states.

SPECIFIC PERFORMANCE CHARACTERISTICS

Prior to release, each lot of ALBAclone[®] Anti-k (cellano) is tested by FDA recommended methods against a panel of antigen-positive and antigen-negative red blood cells to ensure suitable reactivity.

BIBLIOGRAPHY

1. Technical Manual. 15th ed. Bethesda, MD: American Association of Blood Banks, 2005.
2. Standards for Blood Banks and Transfusion Services. 24th ed. Bethesda, MD: American Association of Blood Banks, 2006.

DATE OF ISSUE

April 2009

US Distributor

Quotient Biodiagnostics Inc.
41 University Drive
Newtown
PA 18940
USA

Customer Service Tel: 1-888-284-1901
Product Technical Support Tel: 1-888-228-1990
Customer Service Fax: 1-888-694-5208
E-Mail: customer.service@quotientbiodiagnostics.com
Web: www.quotientbiodiagnostics.com

Manufacturer:

Alba Bioscience Limited
Ellen's Glen Road
Edinburgh
Scotland, UK
EH17 7QT

US License 1807

Tel: +44 (0) 131 536 5907
Fax: +44 (0) 131 536 5897
E-Mail: customer.services@albabioscience.co.uk

© Alba Bioscience Limited 2009

Z137UQPI/01



QUOTIENT
BIODIAGNOSTICS