



## BLOOD GROUPING REAGENT

# Anti-Le<sup>a</sup> ALBAclone® (Murine Monoclonal) For Tube Techniques

REF Z212



### INTERPRETATION OF LABELLING SYMBOLS



Batch code



Use by (YYYY-MM-DD)



Product code



Storage temperature limitation (2-8 °C)



*In vitro* diagnostic medical device



Consult instructions for use

[www.quotientbd.com](http://www.quotientbd.com)



Manufacturer

### INTENDED USE

The Anti-Le<sup>a</sup> reagent is for the *in vitro* detection and identification of human Le<sup>a</sup> positive red blood cells by direct agglutination.

### SUMMARY AND EXPLANATION

Monoclonal Anti-Le<sup>a</sup> and Anti-Le<sup>b</sup> blood grouping reagents enable red blood cells to be classified as one of four phenotypes: Le(a+b-), Le(a-b+), Le(a-b-), Le(a+b+). The latter phenotype, Le(a+b+), is extremely rare.

Agglutination of red blood cells with either of these reagents indicates the presence of the appropriate antigen on the red blood cell surface. Lewis antigens are also present in serum and other body fluids. Cord cells do not express Lewis antigens in sufficient quantity to be agglutinated by these reagents and will therefore group as Le(a-b-). An infant's true Lewis status does not normally become apparent until the age of two years (approximately).

### PRINCIPLE OF THE TEST

When used by the recommended technique, this reagent will cause the agglutination (clumping) of red blood cells carrying the Le<sup>a</sup> antigen. Lack of agglutination demonstrates the absence of the Le<sup>a</sup> antigen.

### REAGENT DESCRIPTION

The main component of this reagent is derived from the *in vitro* culture of the IgM secreting mouse hybridoma:

| Product Name         | Product Code | Cell Line |
|----------------------|--------------|-----------|
| Anti-Le <sup>a</sup> | Z212         | LEA2      |

The formulation also contains bovine serum albumin, potentiator, and 0.1 % (w/v) sodium azide.

NOTE: The volume delivered by the reagent bottle dropper is approximately 40 µL. Care should be taken to ensure that appropriate serum to cell ratios are maintained in all test systems.

This reagent complies with the requirements of Directive 98/79/EC on *in vitro* Diagnostic Medical Devices and the recommendations contained in the Guidelines for Blood Transfusion Services in the United Kingdom.

### STORAGE CONDITIONS

The reagent should be stored at 2-8 °C.

### WARNINGS AND PRECAUTIONS

For *in vitro* diagnostic use only  
Products should be used by qualified personnel  
Do not use beyond the expiration date  
Do not use if turbid  
Do not dilute

The format of the expiration date is expressed as YYYY-MM-DD (Year-Month-Day)

This reagent contains 0.1% sodium azide. Sodium azide may react with lead and copper plumbing to form explosive compounds. If discarded into sink, flush with a large volume of water to prevent azide build-up. Harmful to aquatic life with long lasting effects. Avoid release to the environment. Dispose of contents/container in

accordance with local/regional/national/international regulations.

This reagent is of animal origin (murine and bovine), therefore care must be taken during use and disposal as there is a potential infection risk.

Monoclonal antibodies exhibit a high degree of potency, avidity and specificity. When using such antibodies, great care should be taken to avoid cross contamination.

This product has components (dropper bulbs) containing dry natural rubber.

### SPECIMEN COLLECTION AND PREPARATION

Specimens should be collected by a standard collection technique. The specimen should be tested as soon as possible after collection. If testing is delayed, the specimen should be stored at refrigerated temperatures.

Clotted samples, or those collected in EDTA, should be tested within fourteen days from collection. Donor blood collected in ACD, CPD, CPDA-1, CP2D, CP2D with AS-3, CPD with AS-1, and CPD with AS-5 may be tested until the expiration date of the donation.

Special care should be taken if hemolysed samples must be tested. Grossly icteric or contaminated blood specimens should not be used.

### MATERIALS

#### Material provided

- ALBAclone® Anti-Le<sup>a</sup>

#### Materials required but not provided

- PBS pH 7.0 ± 0.2
- Reagent red blood cells suitable for the control of Anti-Le<sup>a</sup>
- 10 x 75 mm or 12 x 75 mm glass test tubes
- Pipettes
- Centrifuge
- Timer

### PROCEDURE

NOTE: This reagent has been standardised for use by the technique described below and therefore its suitability for use by other techniques cannot be guaranteed.

When a test is required to be incubated for a specific time period, a timer should be used.

It is recommended to allow reagents to reach 18-24 °C prior to use.

When using supplemental testing equipment (i.e. centrifuge), follow the procedures that are contained in the operator's manual provided by the device manufacturer.

### Tube Technique – NIS 15 Minute Incubation/Spin

1. Prepare a 2-3% suspension of red blood cells in PBS pH 7.0 ± 0.2. (Reagent red blood cells may be used directly from the vial or according to the manufacturer's instructions).
2. Add 1 drop of blood grouping reagent to a glass test tube.
3. Add 1 drop of red blood cell suspension. Steps 2 and 3 may be performed in either order.
4. Mix the contents of the test tube and incubate at 18-24 °C for 15 minutes.
5. Centrifuge the test tube.  
NOTE: Suggested centrifugation: 900-1000 g (approx. 3400 rpm) for 10 seconds or a time and speed appropriate for the centrifuge used that produces the strongest reaction of antibody with antigen-positive red blood cells, yet allows easy re-suspension of antigen-negative red blood cells.
6. After centrifugation, gently shake the tube to dislodge the cell button from the bottom and immediately observe macroscopically for agglutination.
7. Record results.

### STABILITY OF REACTION

Test results should be read, interpreted and recorded immediately after centrifugation. Delays may cause dissociation of antigen-antibody complexes resulting in weak positive or false negative reactions.

### INTERPRETATION OF RESULTS

Agglutination = positive test result  
No agglutination = negative test result

### QUALITY CONTROL

Quality control of reagents is essential and should be performed on the day of use.

Le(a+b-) red cells should be used as a positive control.  
Le(a-b+) red cells should be used as a negative control.

### LIMITATIONS

Cord cells do not express Lewis antigens in sufficient quantity to be agglutinated by this reagent and will therefore group as Le(a-b-).

The expression of certain red blood cell antigens may diminish in strength during storage, particularly in EDTA and clotted samples. Better results will be obtained with fresh samples.

Gently re-suspend tube tests before reading. Excessive agitation may disrupt weak agglutination and produce false negative results.

Excessive centrifugation can lead to difficulty in re-suspending the cell button, while inadequate centrifugation may result in agglutinates that are easily dispersed.

False positive or false negative results can occur due to contamination of test materials, improper reaction temperature, improper storage of materials, omission of test reagents and certain disease states.

Suppressed or weak expression of blood group antigens may give rise to false negative reactions.

When using this reagent do not use any optical aid to interpret results.

### SPECIFIC PERFORMANCE CHARACTERISTICS

Prior to release, each lot of ALBAclone® Anti-Le<sup>a</sup> is tested using recommended methods against a panel of antigen-positive and antigen-negative red blood cells to ensure suitable reactivity.

The ALBAclone® Anti-Le<sup>a</sup> reagent reacts with cells expressing the Le<sup>a</sup> antigen.

### BIBLIOGRAPHY

1. British Committee for Standards in Haematology: Guidelines for pre-transfusion compatibility procedures in blood transfusion laboratories, *Trans Med* 2013; 23: 3-35
2. National Blood Service: Guidelines for the Blood Transfusion Services in the United Kingdom, ed 8. TSO, 2013
3. Reid ME, Lomas-Francis C, Olsson ML: The Blood Group Antigen FactsBook, ed 3. Academic Press, 2012

### DATE OF ISSUE

2022-07

For further information or advice please contact your local distributor.



Emergo Europe B.V.  
Prinsessegracht 20  
2514 AP The Hague  
The Netherlands



Alba Bioscience Limited  
James Hamilton Way  
Penicuik  
EH26 0BF  
UK

Tel No: +44 (0) 131 357 3333  
Fax No: +44 (0) 131 445 7125  
E-Mail: [customer.serviceEU@quotientbd.com](mailto:customer.serviceEU@quotientbd.com)  
Web: [www.quotientbd.com](http://www.quotientbd.com)

© Alba Bioscience Limited 2022

Z212PI/08

### INSTRUCTIONS FOR USE

Instructions for Use are available on Quotient website at [www.quotientbd.com](http://www.quotientbd.com) or can be requested from your Local Distributor, by providing the relevant Product Code stated on product labels and Instructions for Use supplied with the product.

### ISTRUZIONI PER L'USO

Le istruzioni per l'uso sono disponibili sul sito Web di Quotient, [www.quotientbd.com](http://www.quotientbd.com), o possono essere richieste al proprio distributore locale fornendo il codice del prodotto indicato sulle etichette del prodotto e nelle istruzioni per l'uso fornite con il prodotto.

### HASZNÁLATI ÚTMUTATÓ

A használati útmutatók elérhetők a Quotient honlapján, [www.quotientbd.com](http://www.quotientbd.com), illetve igényelhetők a helyi forgalmazótól a vonatkozó termék kód megadásával, mely a termék címkéjén és a termékhez mellékelte használati útmutatón található.

### KULLANIM TALÍMATLARI

Kullanım talimatları Quotient web sitesi [www.quotientbd.com](http://www.quotientbd.com) adresinde bulunabilir veya ürün etiketinde veya ürünü birlikte verilen Kullanım Talimatlarında belirtilen ilgili Ürün Kodu sunularak Yerel Distribütörden talep edilebilir.