

BLOOD GROUPING REAGENT

Anti-Lu^b

ALBAclone®

(Murine Monoclonal IgG)
For Tube Technique

REF Z223U

- FOR *IN VITRO* DIAGNOSTIC USE
- Meets FDA potency requirements
- Discard if turbid
- Preservative: 0.1% (w/v) sodium azide

CAUTIONS: THE ABSENCE OF ALL VIRUSES HAS NOT BEEN DETERMINED. THIS PRODUCT HAS COMPONENTS (DROPPER BULBS) CONTAINING DRY NATURAL RUBBER.

INTERPRETATION OF LABELING SYMBOLS

LOT

Batch code



Use by (YYYY-MM-DD)

REF

Product code



Storage temperature limitation (2–8 °C)

IVD

In vitro diagnostic medical device



Consult instructions for use

www.quotientbd.com



Manufacturer

INTENDED USE

The Anti-Lu^b reagent is for the *in vitro* detection and identification of human Lu^b positive red blood cells by direct agglutination.

SUMMARY AND EXPLANATION

Since the description of the antigen Lu^a in 1945 by Callender *et al* and its allele Lu^b in 1956 by Cutbush *et al*, the Lutheran blood group system has been shown to be increasingly complex. Over 10 antigens are now known to be associated with the system and 4 sets of alleles have been identified which are probably controlled from a series of closely linked loci so that Lutheran antigens, like CDE in the Rh system, are inherited as a haplotype.

The antigens of the Lutheran blood group system are not completely developed at birth, show variable strength and are destroyed by trypsin. The low frequency phenotype Lu(a-b-) arises from at least 3 different genetic backgrounds.

PRINCIPLE OF THE TEST

When used by the recommended technique, this reagent will cause agglutination (clumping) of red blood cells carrying the Lu^b antigen. Lack of agglutination demonstrates the absence of the Lu^b antigen.

REAGENT DESCRIPTION

The main component of this reagent is derived from the *in vitro* culture of the mouse hybridoma:

Product Name	Product Code	Cell Line
ALBAclone® Anti-Lu ^b	Z223U	LU2

The formulation also contains bovine material and 0.1% (w/v) sodium azide.

NOTE: The volume delivered by the reagent bottle dropper is approximately 40µL. Care should be taken to ensure that appropriate serum to cell ratios are maintained in all test systems.

STORAGE

The reagent should be stored at 2-8 °C.

WARNINGS AND PRECAUTIONS

For *in vitro* diagnostic use only

Products should be used by qualified personnel

Do not use beyond the expiration date

Do not use if turbid

Do not dilute

The format of the expiration date is expressed as YYYY-MM-DD (Year-Month-Day)

This reagent contains 0.1% (w/v) sodium azide. Sodium azide may be toxic if ingested and may react with lead and copper plumbing to form explosive compounds. If discarded into a sink, flush with a large volume of water to prevent azide buildup. This reagent is of animal origin (murine and bovine), therefore care must be taken during use and disposal as there is a potential infection risk.

CAUTION: SOURCE MATERIAL FROM WHICH THIS PRODUCT WAS DERIVED, WAS FOUND NEGATIVE FOR INFECTIOUS AGENTS WHEN TESTED IN ACCORDANCE WITH CURRENT FDA REQUIRED TESTS.

The bovine material which was used has been collected in a USDA approved facility.

Monoclonal antibodies exhibit a high degree of potency, avidity and specificity. When using such antibodies, great care should be taken to avoid cross contamination.

This product has components (dropper bulbs) containing dry natural rubber.

SPECIMEN COLLECTION AND PREPARATION

Specimens should be collected by a standard collection technique. The specimen should be tested as soon as possible after collection. If testing is delayed, the specimen should be stored at refrigerated temperatures.

Clotted samples, or those collected in EDTA, should be tested within fourteen days from collection. Blood collected into other anticoagulants may be used (ACD, CPD and ACD A1). Donor blood may be tested until the expiration date of the donation.

Special care should be taken if hemolyzed samples must be tested. Grossly icteric or contaminated blood specimens should not be used.

MATERIALS

Material provided

- ALBAclone® Anti-Lu^b

Materials required but not provided

- Isotonic saline
- Reagent red blood cells suitable for the control of Anti-Lu^b
- 10 x 75mm or 12 x 75mm glass test tubes
- Pipettes
- Optical aid (optional)
- Centrifuge
- Timer
- Heating block/waterbath

PROCEDURE

NOTE: This reagent has been standardized for use by the technique described below and therefore its suitability for use in other techniques cannot be guaranteed. When a test is required to be incubated for a specific time period, a timer should be used.

When using supplemental testing equipment (i.e. centrifuge), follow the procedures that are contained in the operator's manual provided by the device manufacturer.

Tube Technique - 15 Minute Incubation/Spin

- Add 1 drop of blood grouping reagent to a glass test tube.
- Add 1 drop of red blood cells suspended to 2-4% in isotonic saline. Reagent red cells may be used as provided (preservative suspended).
- Mix the contents of the test tube well and incubate at 37 °C ± 1 °C for 15 minutes.
- Centrifuge the test tube.

NOTE: Suggested centrifugation: 900-1000g (approx. 3400 rpm) for 10 seconds or a time and speed appropriate for the centrifuge used that produces the strongest reaction of

antibody with antigen-positive red blood cells, yet allows easy resuspension of antigen-negative red blood cells.

5. After centrifugation, gently shake the tube to dislodge the cell button from the bottom and immediately observe macroscopically for agglutination. Negative reactions may be examined with an optical aid.
6. Record results.

STABILITY OF REACTION

Test results should be read, interpreted and recorded immediately after centrifugation. Delays may cause dissociation of antigen-antibody complexes resulting in weak positive or false negative reactions.

INTERPRETATION OF RESULTS

Agglutination	=	positive test result
No agglutination	=	negative test result

QUALITY CONTROL

Quality control of reagents is essential and should be performed on each day of use and in accordance with local, state and federal regulations.

Lu(a+b+) red blood cells should be used as a positive control
Lu(a+b-) red blood cells should be used as a negative control

LIMITATIONS

Heating blocks and waterbaths promote better heat transfer and are recommended for 37 °C tests, particularly where the incubation period is 30 minutes or less.

The expression of certain red blood cell antigens may diminish in strength during storage, particularly in EDTA and clotted samples. Better results will be obtained with fresh samples.

Gently re-suspend tube tests before reading. Excessive agitation may disrupt weak agglutination and produce false negative results.

Excessive centrifugation can lead to difficulty in re-suspending the cell button, while inadequate centrifugation may result in agglutinates that are easily dispersed.

False positive or false negative results can occur due to contamination of test materials, improper reaction temperature, improper storage of materials, omission of test reagents and certain disease states.

Suppressed or weak expression of blood group antigens may give rise to false negative reactions.

SPECIFIC PERFORMANCE CHARACTERISTICS

Prior to release, each lot of ALBAclone® Anti-Lu^b is tested using FDA recommended methods against a panel of antigen-positive and antigen-negative red blood cells to ensure suitable reactivity.

This Anti-Lu^b reagent will give significantly weaker reactions with red blood cells of the phenotypes Lu(a-b⁺) and Lu(a+b⁺).

BIBLIOGRAPHY

1. Roback, JD, Grossman BJ, Harris T, *et al.* AABB Technical Manual, 18th ed. AABB, 2014.
2. AABB Standards Program Committee. Standards for Blood Banks and Transfusion Services, 29th ed. AABB 2014.
3. Reid ME, Lomas-Francis C, Olsson ML: The Blood Group Antigen FactsBook, ed 3. Academic Press, 2012.

DATE OF ISSUE

2019-07-24

US Distributor

Quotient
301 South State Street
S-204
Newtown
PA 18940
USA

Customer Service Tel: 1-888-284-1901
Product Technical Support Tel: 1-888-228-1990
Customer Service Fax: 1-888-694-5208
E-Mail: customer.serviceUS@quotientbd.com
Web: www.quotientbd.com/us



Alba Bioscience Limited
James Hamilton Way,
Penicuik, EH26 0BF,
UK

U.S. License 1807